

ADDENDUM NO. 1

Tender Package Name Supply and Distribution of Male Latex Condom
Tender Publication Date 16 January, 2018
Invitation Ref. No. icddr,b/SCM/OTM/2018/01 dated 16 Jan 2018
Published In "The Daily Prothom Alo" and "The Daily Star"

Section Number	Originally Provided	To be Substituted By	Remarks
Section VII: Technical Specifications Page no. 60-67	All condoms supplied under this invitation are to be in accordance with the WHO/UNFPA specifications, 2010.	All condoms supplied under this invitation are to be in accordance with the WHO/UNFPA specifications, 2010 (Revised 2013), updated on July 2017.	Bidders are advised to follow the WHO/UNFPA 2010 (revised 2013) version of the specification for male latex condom for guidance. Specific requirement is given in Section VII. Technical Specifications, Page no. 60-67.
Section VII. Technical Specifications, Page no. 62, Sl. no. 3.7, Length	Requirement : 170mm to 180mm	Requirement : 165mm to 180mm	This shall be replaced in all corresponding clauses, sub-clauses and relevant sections of the bid document, where applicable.


25.1.18



Gayatri Khan Mahub
Senior Manager, Procurement
icddr,b, Mohakhali, Dhaka

